





## Microphone Frequency Response Measurement Report

**Microphone Manufacturer:** Microtech Gefell    **Model Number:** M940  
**Serial Number:** 0146    **Phantom Power:** 48V


**Measurement Date:** August 06, 2021    **Temperature:** 79°F/26°C  
**Technician:** P. Campbell    **Humidity:** 50%



1/3 OB Results	
Freq.	Resp. (dB)
20	0.57
25	0.49
31.5	0.39
40	0.25
50	0.09
63	-0.12
80	-0.35
100	-0.54
125	-0.63
160	-0.62
200	-0.59
250	-0.56
315	-0.51
400	-0.46
500	-0.39
630	-0.29
800	-0.17
1000	0.00
1250	0.16
1600	0.23
2000	0.01
2500	-0.41
3150	-0.48
4000	-0.48
5000	-0.56
6300	0.73
8000	0.81
10000	-0.46
12500	0.51
16000	0.93
20000	-2.63
25000	-12.79

This frequency response measurement is not a NIST-traceable calibration.

### Reference Microphone





## Microphone Frequency Response Measurement Report

### Supplemental Information

<b>Microphone Manufacturer:</b>	Microtech Gefell	<b>Model Number:</b>	M940
<b>Serial Number:</b>	0146	<b>Phantom Power:</b>	48V
<b>Measurement Date:</b>	August 06, 2021	<b>Temperature:</b>	79°F/26°C
<b>Technician</b>	P. Campbell	<b>Humidity:</b>	50%
<b>Ref. Calibrator:</b>	Larson Davis CAL200		
<b>Serial Number:</b>	8477		
<b>Microphone Sensitivity:</b>	37 dB (95 mV/Pa)	<b>Noise Floor:</b>	10 dBA
<b>(re 1000Hz, 94 dB)</b>			

