



Left input to the QA401 is shorted with a 50 Ohms BNC terminator plug

Right channel is loopback test

The generator noise is the difference between the two - close to 20 dB